



Electrotechnical Research and Projecting Institute j.s.c.,
SKTC 101 Nová Dubnica, Slovakia

Notified Body/Competent Body
Identification No. 1293



CERTIFICATE OF CONFORMITY

No. 40019/101/1/2005/CE

Manufacturer
L.C. MANAGER, s.r.o.
Župčany 176, 080 01 Prešov, Slovakia

Product
Rubbered bed sheet for incontinent patients

Types
AMBITUR 715, AMBITUR 880, PROTEC 16S

Classification
Medical Device Class I

The sample meets the essential safety requirements of the following directives:
Medical Device Directive 93/42/EEC

A sample of the product was tested and found to be in conformity with following standards and/or technical specifications applied:

ENV 14237:2002
EN 1413:1998
EN ISO 14184-1:1998
STN 80 0055: 2004
EN ISO 6330:2000
EN 25077:1993


Test Report No. 40019/2005

This certificate is the result of test carried out on one sample and does not represent the serial production of this product.

Valid from: June 10th, 2005
Valid until: June 09th, 2010

Nová Dubnica, June 10th, 2005

EVPÚ, a.s., SKTC 101
Trenčianska 19,
018 51 Nová Dubnica
Slovakia
377923



Signature

Tel.: ++421 (0)42 4403 600, 4403 500
Fax: ++421 (0)42 4403 502
e-mail: sktc101@evpu.sk